



**NATIONAL INSTITUTE OF TECHNOLOGY
ROURKELA-769008(ODISHA)**
An Institute of National Importance under Ministry of Education, GOI

NOTICE INVITING TENDER

Tender Notification No: NITR/PW/EE/2021/245

Dated: 22.02.2021

The National Institute of Technology, Rourkela invites bids from the eligible bidders for the service for Calibration of High Voltage Laboratory Equipments for NABL Accreditation.at Dept. of Electrical Engineering, NIT Rourkela.

Last date of Submission of Bid : 05/03/2021 by 03:00 PM

Opening date of Techno-commercial & financial bid : 08/03/2021 at 11:00 AM

For Details: [_ https://nitrkl.ac.in/OldWebsite/Jobs_Tenders/8Maintenance/Default.aspx](https://nitrkl.ac.in/OldWebsite/Jobs_Tenders/8Maintenance/Default.aspx)

Contact:

**Dr. Subrata Karmakar [EE];
Department of Electrical Engg.
Ph. 0661-2462411;
Email: karmakars@nitrkl.ac.in**

Bidding through: <https://eprocure.gov.in/eprocure/app>

**Sd/-
Registrar**



NATIONAL INSTITUTE OF TECHNOLOGY ROURKELA – 769 008, ODISHA

(Single TENDER NOTICE NO: NITR/PW/EE/2021/245

Date: 22.02.2021)

Calibration of High Voltage Laboratory Equipments for NABL Accreditation

Sl.No	Description of Goods/Service	Quantity																
1.	<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th colspan="4" style="text-align: left;">500kV AC Power Frequency Test Setup.</th> </tr> <tr> <th style="width: 10%;">SL No.</th> <th style="width: 35%;">Name of Equipment</th> <th style="width: 20%;">Model/Type</th> <th style="width: 35%;">Range/Ratings</th> </tr> </thead> <tbody> <tr> <td style="text-align: center;">1</td> <td>Power Frequency High Voltage Testing Transformer cascaded of two 250kVA transformer with Coupling Capacitor and Measuring Capacitor and a Sphere Gap.</td> <td>Make: WSTS Test Systems, Type-TEOK 500-500</td> <td>0.6/500kV, 500 kVA, 2 Phase, 50 Hz PD Level <5pC</td> </tr> </tbody> </table>	500kV AC Power Frequency Test Setup.				SL No.	Name of Equipment	Model/Type	Range/Ratings	1	Power Frequency High Voltage Testing Transformer cascaded of two 250kVA transformer with Coupling Capacitor and Measuring Capacitor and a Sphere Gap.	Make: WSTS Test Systems, Type-TEOK 500-500	0.6/500kV, 500 kVA, 2 Phase, 50 Hz PD Level <5pC	1.00				
500kV AC Power Frequency Test Setup.																		
SL No.	Name of Equipment	Model/Type	Range/Ratings															
1	Power Frequency High Voltage Testing Transformer cascaded of two 250kVA transformer with Coupling Capacitor and Measuring Capacitor and a Sphere Gap.	Make: WSTS Test Systems, Type-TEOK 500-500	0.6/500kV, 500 kVA, 2 Phase, 50 Hz PD Level <5pC															
2.	<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th colspan="4" style="text-align: left;">400kV Impulse Voltage Test Setup.</th> </tr> <tr> <th style="width: 10%;">SL No.</th> <th style="width: 25%;">Name of Equipment</th> <th style="width: 20%;">Model/Type</th> <th style="width: 45%;">Range/Ratings</th> </tr> </thead> <tbody> <tr> <td style="text-align: center;">1</td> <td>Impulse Voltage Generator</td> <td>Make:Haefly, USA Type:33</td> <td>400kV, 20 kJ, 2 Stage with output of either +ve or -ve polarity Output wave shape 1.2 / 50 μs with permissible tolerance as per IS 2071 and Output wave shape 250 / 2500 μs with permissible tolerance as per IS 2071</td> </tr> <tr> <td style="text-align: center;">1.a</td> <td>Peak Voltmeter</td> <td>Make: Haefly, USA Type-65</td> <td>200/400/800/1600 V</td> </tr> </tbody> </table>	400kV Impulse Voltage Test Setup.				SL No.	Name of Equipment	Model/Type	Range/Ratings	1	Impulse Voltage Generator	Make:Haefly, USA Type:33	400kV, 20 kJ, 2 Stage with output of either +ve or -ve polarity Output wave shape 1.2 / 50 μ s with permissible tolerance as per IS 2071 and Output wave shape 250 / 2500 μ s with permissible tolerance as per IS 2071	1.a	Peak Voltmeter	Make: Haefly, USA Type-65	200/400/800/1600 V	1.00
400kV Impulse Voltage Test Setup.																		
SL No.	Name of Equipment	Model/Type	Range/Ratings															
1	Impulse Voltage Generator	Make:Haefly, USA Type:33	400kV, 20 kJ, 2 Stage with output of either +ve or -ve polarity Output wave shape 1.2 / 50 μ s with permissible tolerance as per IS 2071 and Output wave shape 250 / 2500 μ s with permissible tolerance as per IS 2071															
1.a	Peak Voltmeter	Make: Haefly, USA Type-65	200/400/800/1600 V															
3.	<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th colspan="4" style="text-align: left;">100kV AC power frequency test Setup for partial discharge (PD) measurement.</th> </tr> <tr> <th style="width: 10%;">SL No.</th> <th style="width: 35%;">Name of Equipment</th> <th style="width: 20%;">Model/Type</th> <th style="width: 35%;">Range/Ratings</th> </tr> </thead> <tbody> <tr> <td style="text-align: center;">1</td> <td>Power Frequency High Voltage Testing Transformer with Coupling Capacitor, Measuring Capacitor, PD Meter, PD pulse generator and doubled shelled room.</td> <td>Make: WSTS Test Systems, Type-TEO 100/20</td> <td>0.22/100kV, 10 kVA, 0.5 A, 1 Phase, 50 Hz PD Level <5pC</td> </tr> </tbody> </table>	100kV AC power frequency test Setup for partial discharge (PD) measurement.				SL No.	Name of Equipment	Model/Type	Range/Ratings	1	Power Frequency High Voltage Testing Transformer with Coupling Capacitor, Measuring Capacitor, PD Meter, PD pulse generator and doubled shelled room.	Make: WSTS Test Systems, Type-TEO 100/20	0.22/100kV, 10 kVA, 0.5 A, 1 Phase, 50 Hz PD Level <5pC	1.00				
100kV AC power frequency test Setup for partial discharge (PD) measurement.																		
SL No.	Name of Equipment	Model/Type	Range/Ratings															
1	Power Frequency High Voltage Testing Transformer with Coupling Capacitor, Measuring Capacitor, PD Meter, PD pulse generator and doubled shelled room.	Make: WSTS Test Systems, Type-TEO 100/20	0.22/100kV, 10 kVA, 0.5 A, 1 Phase, 50 Hz PD Level <5pC															

Note : 1) Organisation/Bidder Should be NABL Certified Laboratory 2) Price including fooding/ Lodging /installation etc. 3) Aforesaid three items will includes all prices.

1. **Quantity required Delivery:** As mentioned above (All information provided in technical specification)
2. **Contract Period:** Within 365 days from the date of purchase order
3. **Last Date of submission of Tender** : 05.03.2021 by 03: 00 PM
4. **Date of opening of Technical and Financial Bid** : 08.03.2021 at 11:00 AM
5. The firm should not have been black listed at any time.

6. The submission of following bids by the tenderer should be through <https://eprocure.gov.in/eprocure/app> Please follow the guidelines as per the portal.

**Calibration of High Voltage Laboratory Equipments for NABL Accreditation
(SingleTender Notice No.: - NITR/PW/EE/2021/245, dated: 22.02.2021
Due on 05.03.2021 at 03:00 PM)**

7. Liquidated damage clause will be charged for any delay in supply of goods.
8. The validity of the tender shall be **90 days** from the date of opening of techno-commercial & financial bid.
9. Detailed advertisement including all tender documents is also available in our website at https://nitrkl.ac.in/OldWebsite/Jobs_Tenders/8Maintenance/Default.aspx
10. NIT reserves the right to qualify or deny prequalification of any or all applicants without assigning any reasons.

(REGISTRAR)
NIT, Rourkela
Fax No- 0661-2462022
Ph. No -0661-2472021